

**Miller, Diane M.**

---

**From:** Schorer Bruce  
**Sent:** Thursday, November 28, 2002 6:39 AM  
**To:** NIOCINDOCKET@CDC.GOV  
**Subject:** Standards Development for APR ESCAPE Respirators

The applicable 42 CFR PART 84 "APPLICABLE SECTIONS" .....

(9) reads :- .....Service and Maintenance

(9) should read.....*Routine Inspection* , Service and Maintenance

Comment : Routine Inspection is recommended to capture the need to service & maintain these standby products in a "Proactive manner"

<<Bruce Schorer>>